



MISSOURI DEPARTMENT OF HEALTH  
STATE PUBLIC HEALTH LABORATORY  
**DATAMASTER MAINTENANCE REPORT**

PATH / ALCOHOL PROGRAM

JUL-87

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file. **LESS \$1,000.00**

DATAMASTER SN **204103** DATE OF INSPECTION **5-27-2009**

LOCATION OF INSTRUMENT (STREET AND CITY) **SCHUYLER COUNTY SHERIFF'S DEPT. LANCASTER, MO.** TIME OF INSPECTION **1021**

**CHECKLIST:** Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER **50** °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C) **34°c**

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 **0.099**

TEST 2 **0.100**

TEST 3 **0.101**

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS **0** (0-.04) **0** (.05-.09) **0** (.10-.14) **0** (.15-.19) **1** (Over .19) **1**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

**OPERATES WITHIN DOH GUIDELINES**

**LOT # 08340**

**GUTH LABORATORIES, INC.**

**EXP: 10/15/2009**

**VAPOR CONCENTRATION .10%**

**INSPECTING OFFICER**

SIGNATURE

**TPR. Adam M. Kinney, 582**

PRINT NAME

**ADAM M. KINNEY**

TYPE II PERMIT NUMBER/EXPIRATION DATE

**820062 2/25/2010**

TELEPHONE NUMBER

**(660) 385-2132**



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **08340** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography and found to contain **0.1211** percent (w/vol) ethyl alcohol. The expiration date for this lot number is **October 15, 2009** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.10** percent BAC.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

Face This Side Down - This Edge In First

## BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 284103  
05/27/09  
18:21

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
HEATERS:	
SAMPLE CHAMBER:	50:
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdeefghijklmnop  
qrstuvwxyz{|}~

Operator Signature

*TR. Adam M. King, 581*

Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

## BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204103  
05/27/09

ARREST TIME: 10:06  
SUBJECT NAME:  
SAMPLE A/B  
DOB: 05/27/77 SEX: M  
STATE/D.L.: MO/123456789  
ARRESTING OFFICER:  
KINNEY/ADAM/M  
OFFICER I.D.: 581  
TESTING OFFICER:  
KINNEY/ADAM/M  
OFFICER I.D.: 581  
PERMIT NUMBER: 820062  
EXPIRATION DATE: 02/25/10  
MISCELLANEOUS DATA:  
.10 RFI

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

*TP. Adam M. Kinney, 581*

Printed on recycled paper with agri-based inks

CMSU

Face This Side Down - This Edge In First

## BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204103  
05/27/09

TESTING OFFICER:  
KINNEY/ADAM/M  
OFFICER I.D.: 581  
PERMIT NUMBER: 820062  
EXPIRATION DATE: 02/25/10  
MISCELLANEOUS DATA:  
.10 MAINT.

--- SUPERVISOR MODE ---

BLANK TEST	.006	10:23
INTERNAL STANDARD	VERIFIED	10:23
EXTERNAL STANDARD	.099	10:23
BLANK TEST	.006	10:24
EXTERNAL STANDARD	.100	10:24
BLANK TEST	.006	10:25
EXTERNAL STANDARD	.101	10:25
BLANK TEST	.006	10:26

M = 3  
SIM. = .1  
AVG. = .1

Operator Signature

*TP. Adam M. Kinney, 581*

Printed on recycled paper with agri-based inks

CMSU 2208-02

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



ADAM KINNEY



Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/25/08  
Number 820062  
Expires 02/25/2010

MO 580-0771 (7-88)

  
Director of State Public Health Laboratory  
  
Director, Department of Health

Lab. 4 (R7-88)